

Test Results of the Effects of Silver Antimicrobial Dressings on Drug Resistant Bacteria *in vitro*

Brand name	Product types available	Log reduction in 30 minutes for MRSA	Log reduction in 30 minutes for VRE	Log reduction in 30 minutes for 5 NDM-1 producing bacterial strains	Log reduction in 30 minutes for drug resistant <i>P. aeruginosa</i> 137366	Method of silver delivery		What does that mean?
						Initial	Solution (concentration)	
ACTICOAT [◇] Antimicrobial Barrier Dressing (Smith+Nephew)	Dressing (3 and 7 day)	>6.0 ¹	>6.0 ¹	>4.0 ⁷	>5.0 ⁸	Silver Ag ⁰ Nanocrystals ¹⁰	Ag ⁺ , Ag ⁰ / Ag ⁺ clusters (70-100 ppm) ¹⁰	ACTICOAT dressing showed marked bactericidal activity against MRSA, VRE, NDM-1 producing bacterial strains and drug-resistant <i>P. aeruginosa</i> at 30 minutes (<i>in vitro</i>)
Silverlon TM (Argentum LLC)	Dressing, island dressing, foam strips (7 day)	0.0 ¹	1.54 ¹	Not reported	<1 ⁸	Silver oxide ¹¹	Ag ⁺ (<1 ppm) ¹¹	Silverlon showed no activity against MRSA and bacteriostatic activity against VRE within 30 minutes. No published data to support rapid antimicrobial activity against NDM-1 producing bacterial strains or drug-resistant <i>P. aeruginosa</i>
AQUACEL TM Ag (ConvaTec)	Dressing, rope (7 Day)	0.0 ¹	0.0 ¹	Not reported	<1 ⁸	AgCl ¹¹	Ag ⁺ (<1-3 ppm) ^{1a}	Aquacel Ag showed no activity against either MRSA or VRE within 30 minutes. No published data to support rapid antimicrobial activity against NDM-1 producing bacterial strains or drug-resistant <i>P. aeruginosa</i>
Aquacel TM Ag Extra TM (ConvaTec)	Dressing (7 Days)	<1 ²	<1 ²	Not reported	0.0 ²	AgCl ¹¹	Ag ⁺ (<1-3 ppm) ¹	Aquacel Ag Extra showed no activity against either MRSA or VRE within 30 minutes. No published data to support rapid antimicrobial activity against NDM-1 producing bacterial strains or drug-resistant <i>P. aeruginosa</i>
TheraBond 3D TM	Wound layer, island, wrap dressing(7 day)	0.0 ³	No Speed of kill data	Not reported	<1 ³	Silver plated fabric	Ag ⁺ Not reported	Therabond 3D demonstrated bactericidal activity against MRSA, but no rapid activity against VRE was reported. No published data to support rapid antimicrobial activity against NDM-1 producing bacterial strains or drug-resistant <i>P. aeruginosa</i>
KerraContact TM (Crawford Healthcare)	Barrier dressing (7 day)	0.0 ⁴	0.0 ⁴	Not reported	Not reported	Silveroxysalts ¹²	Ag ⁺ , Ag ²⁺ , Ag ³⁺ (6 ppm) ^{12,13}	KerraContact is bactericidal against MRSA or VRE from 4 hours with no effect at 30 minutes reported. No published data to support rapid antimicrobial activity against NDM-1 producing bacterial strains or drug-resistant <i>P. aeruginosa</i>
Mepilex TM Ag (Mölnlycke)	Foam, post-op border, transfer (7 day)	<1 ⁵	<1 ⁶	Not reported	<1 ⁹	Silver sulphate ¹⁴	Ag ⁺ (6 ppm) ^{15,16}	Mepilex Ag showed no activity against MRSA or VRE in 30 minutes. No published data to support rapid antimicrobial activity against NDM-1 producing bacterial strains or drug-resistant <i>P. aeruginosa</i>
Silvercel TM (Systagenix)	Dressing, rope (7 day)	<1 ⁶	0.0 ⁶	Not reported	Not reported	Ag alginate ¹⁷	Ag ⁺ (~1 ppm) ¹¹	Silvercel showed no activity against MRSA or VRE in 30 minutes. No published data to support rapid antimicrobial activity against NDM-1 producing bacterial strains or drug-resistant <i>P. aeruginosa</i>
SilvaSorb TM (Medline Ind.)	Sheet, perforated sheet, strands, (7 day), gel(3 day)	0.0 ¹	0.0 ¹	Not reported	<1 ⁸	AgCl ¹¹	Ag ⁺ (~1 ppm) ¹¹	SilvaSorb showed no activity against MRSA or VRE in 30 minutes No published data to support rapid antimicrobial activity against NDM-1 producing bacterial strains or drug-resistant <i>P. aeruginosa</i>

The Minimum Bactericidal Concentration (MBC) for clinically relevant bacteria is between 5 and >50ppm^{18,19} and up to 60.5ppm for MRSA²⁰. This means in order to kill MRSA the dressing must kill bacteria with a MBC of at least 60.5 PPM (200ppm x 0.0325=60.5ppm). Notes: In vitro results do not necessarily translate into clinically relevant results. Antimicrobial agents are considered bactericidal if they induce a 3 or greater log reduction in a clinically relevant timeframe¹⁹ These results do not represent a head-to-head safety/efficacy clinical comparison.

Test Results of the Effects of Silver Antimicrobial Dressings on Drug Resistant Bacteria *in vitro*

For detailed product information, including indications for use, contraindications, effects, warnings and precautions, please consult the product's Instructions for Use (IFU) prior to use.

References

1. Data on file: Testing performed by Hoosier Microbiological Laboratory, Muncie, IN (independent lab). January 3, 2005.
2. Data on file: 1405007. Antimicrobial activity of DURAFIBER™, DURAFIBER Ag, Aquacel™ Extra, Aquacel Ag Extra and Aquacel Ag+ Extra against bacteria, yeast and fungi commonly found in wounds over a 7 day period.
3. Data on file: 0911020. Antimicrobial assessment of ACTICOAT™ Flex 7 in comparison to Therabond™ against MRSA and Pseudomonas aeruginosa, 2009.
4. Exsalt™ SD7-4 Brochure. PR-0045.0.5.Data on file: 070805
5. Antimicrobial activity of ALLEVYN™ Ag Dressings in comparison to other silver-based antimicrobial dressings against MRSA using a dynamicshake flask method, 2007.
6. Data on file: Testing performed by Hoosier Microbiological Laboratory, Muncie, IN (independent lab). Raw data/XL file.
7. Hope R, et al. The in vitro antibacterial activity of Nanocrystalline Silver dressings against bacteria with NDM-1 carbapenemase [poster]. EWMA; 2012.
8. Data on file: Testing performed by Hoosier Microbiological Laboratory, Muncie, IN (independent lab). January 4, 2005.
9. Data on file: 0707051. Antimicrobial activity of ALLEVYN Ag Dressings in comparison to other silver-based antimicrobial dressings against Pseudomonas aeruginosausing a dynamic shake flask method, 2007.
10. Dunn K, et al. The role of ACTICOAT with Nanocrystalline Silver in the management of burns. Burns. 2004; 30 Suppl 1:S1-9.
11. Ovington LG. Why is Nanocrystalline Silver Superior? WOUNDS. 2004; 13(2):1-9.
12. Exsalt SD7-4 Brochure. PR-0046.0.
13. Djokić N, et al. Coating of High Density Polyethylene (HDPE) with Silver Compounds and Their Antimicrobial Activity in vitro. ECS Transactions. 2008; 11(21):59-68.
14. Mepilex™ Ag 510k (K071354).
15. Schwarzkopf A, et al. "The Release of Silver Ions from Foam Dressings". Wound Management. 2010; 4(1): 1-4.
16. Mepilex Ag European Patent Application: 07004275.9.
17. Silvercel™ Brochure. AWC-SC001.
18. Yin, H.Q. et al. Comparative Evaluation of the Antimicrobial Activity of ACTICOAT Antimicrobial Barrier Dressing. J Burn Care Rehabil 1999;20:195-200.
19. Hall, R.E. et al. Inhibitory and Cidal Antimicrobial Actions of Electrically Generated Silver Ions. J Oral Maxillofac Surg 1987;45:779-784.
20. Maple, P.A. et al. Comparison of the in vitro activities of the topical antimicrobials azelaic acid, nitrofurazone, silver sulphadiazine and mupirocin against methicillin-resistant Staphylococcus aureus. Journal of Antimicrobial Chemotherapy. 1992;29:661-668.